

Instruction Sheet for the AWS 200 ml Attachment HEMC PN 000A3393

Warning: This attachment is made of **GLASS**. It is fragile. Breakage is **NOT** covered under our warranty.

This attachment will allow you to double the maximum tidal volume (TV) deliverable from your Anesthesia WorkStation.

Removal of the currently installed ventilator tube assembly:

Carefully remove the 100 ml ventilator tube assembly by loosening the two thumbscrews that hold the tube cap to the upper manifold. When the cap is loose, **GENTLY** lift the vent tube off the mounting nipple on the main manifold at the bottom. **Take care NOT to tilt the glass tube too far before it is clear of the nipple. This may result in tube breakage.**

As you slide the vent tube off the nipple, keep the piston within the tube. Avoid touching the sides of the piston or the inside of the glass tube or wear surgical gloves to prevent finger oils from getting on these surfaces.

When the assembly is free, remove the piece of foam from within the 200 ml Attachment and insert it into the open end of the cylinder being removed such that the piston will not slide out.

Installation of the new 200 ml attachment:

Remove bottom bushing from tube. Place the bushing onto the main manifold mounting nipple.

Without allowing the piston to slide out, place the new attachment onto the mounting nipple bushing of the main manifold. Again, use a twisting motion and be sure the bushing in the end of the tube remains flush with the bottom of the tube and that it does not ride up into the tube.

Rotate the tube to orient it such that the scale is facing forward.

Adjust the 200mL Vent Tube Cap to align thumbscrews with top manifold. Once aligned tighten thumbscrews into place.

Verify that the piston moves freely up and down the length of the tube by filling it with gas then letting it out and allowing the piston to drift down to the bottom.

Cleaning and Maintenance:

Refer to the Cleaning and Maintenance section of the AWS Operation Manual.