

4.2. MRI Safety information

The Ambu Disposable PEEP 20 Valve is tested to be MR Conditional and therefore may be safely used in the MR environment (not inside the MR bore) under the following conditions.

- Static magnetic field of 7 Tesla and less, with
- Maximum spatial field gradient of 16,000 G/cm (160 T/m)
- Maximum force product of 721,000,000 G²/cm (721 T²/m)

Use inside the MR bore may influence MR image quality.

RF-induced heating and MR image artifacts have not been tested.

Any metallic parts are fully encapsulated and do not have any contact with the human body.

4.2. MRI Safety information

The Ambu Disposable PEEP 20 Valve is tested to be MR Conditional and therefore may be safely used in the MR environment (not inside the MR bore) under the following conditions.

- Static magnetic field of 7 Tesla and less, with
- Maximum spatial field gradient of 16,000 G/cm (160 T/m)
- Maximum force product of 721,000,000 G²/cm (721 T²/m)

Use inside the MR bore may influence MR image quality.

RF-induced heating and MR image artifacts have not been tested.

Any metallic parts are fully encapsulated and do not have any contact with the human body.